



(For scientific research use only, not for clinical diagnosis!)

Swine influenza virus antibody (EIV-Ab)

ELISA kit instruction manual Product

number: BY-EP771477 Specifications:

48T/96T

**Please read the instructions carefully before use. If you have any questions,
please contact us through the following methods: Official hotline: 025-5229-
8998 Sales department phone: 13914481711 Technical phone: 15950492658
Contact email: 3224949330@qq.com Company website:
www.byabscience.cn For specific shelf life, please see the reagents Box
packaging label. Please use the kit within the shelf life.**

When contacting us, please provide the product number and production date (see box label) so that we can serve you more efficiently.

Nanjing BYabscience technology Co.,Ltd

Website: www.byabscience.cn

Official hotline: 025-5229-8998

Supervision phone number:



Kit performance Physical properties: Each liquid component is clear and transparent, with no precipitation or floc. Microplate aluminum foil bags should be vacuum packed without damage or leakage.

Negative control OD value: less than 0.2.

Positive control OD value: greater than 0.8.

Precision: intra-batch variation coefficient CV% is less than 10%; inter-batch variation coefficient CV% is less than 15%.

Recovery rate: The recovery rate is between 85%-115%.

Specificity: This kit recognizes natural and recombinant swine influenza virus antibodies (EIV-Ab) and has no crossover with structural analogs. **Stability:** Stored at 2°C-8°C, validity period is 6 months.

Purpose: For qualitative detection of swine influenza virus antibodies (EIV-Ab) in samples such as serum, plasma, cell culture supernatant, and tissue.

Shelf life: Stored at 2°C-8°C, valid for 6 months.

Experimental principle

The kit uses an indirect enzyme-linked immunosorbent assay (ELISA). To the microwells pre-coated with swine influenza virus antibody (EIV-Ab) capture antigen, add the specimen, negative and positive controls in sequence, then add the HRP-labeled detection antibody, incubate and wash thoroughly. The substrate TMB is used for color development. TMB is converted into blue under the catalysis of peroxidase and into the final yellow under the action of acid. The depth of the color is positively correlated with the swine influenza virus antibodies (EIV-Ab) in the sample. Use a microplate

reader to measure the absorbance (OD value) at a wavelength of 450 nm to determine negative and positive.

Nanjing BYabscience technology Co.,Ltd

Website: www.byabscience.cn

Official hotline: 025-5229-8998

Supervision phone number:



Kit components and storage: Store unopened kits at 2-8 degrees Celsius. Do not use expired kits.

| Components | 48-well configuration | 96-well configuration | Store after opening |
|-------------------------|-----------------------|-----------------------|----------------------|
| Pre-coated enzyme plate | 48T | 96T | 2-8°C 14 days |
| negative control | 0.3mL | 0.3mL | 2-8°C 14 days |
| positive control | 0.3mL | 0.3mL | 2-8°C 14 days |
| sample diluent | 3ml | 6ml | 2-8°C 180 days |
| HRP labeled antibodies | 5ml | 10ml | 2-8°C 14 days |
| Chromogenic substrate A | 3ml | 6ml | 2-8°C 180 days |
| Chromogenic substrate B | 3ml | 6ml | 2-8°C 180 days |
| stop solution | 3ml | 6ml | 2-8°C 180 days |
| 20×Lotion | 15ml | 25ml | 2-8°C 180 days |
| sealing film | 2 sheets | 2 sheets | |
| manual | 1 serving | 1 serving | |
| Ziplock bag | 1 | 1 | |

Note: 1: Please check whether the label and quantity of the reagents in the kit are consistent with the table before use.

2: If the components of the kit need to be used again, please ensure that they have not been contaminated since the last use. 3: If the enzyme plate is not used up in a single time, remember to seal it and store it at 2-8°C.

Prepare your own test equipment required for the test (not provided, but can assist in

1) Microplate reader capable of detecting absorbance at 450 nm 2) Pipette, pipette tip, and sample addition tank 3) 37°C incubator or water bath 4) Test tubes, centrifuge tubes, measuring cylinders, etc. for preparing reagents 5) Distilled water or deionized water Ionized water 6) Vortex oscillator, microplate oscillator

Nanjing BYabscience technology Co.,Ltd

Website: www.byabscience.cn

Official hotline: 025-5229-8998

Supervision phone number:



**Notes 1) For scientific research use only,
not for clinical diagnosis.**

- 2) Use within the validity period marked on the kit. Expired products must not be used.
- 3) Do not mix with kits or components from other manufacturers. Use the sample diluent provided with the kit.
- 4) If the sample value is higher than the highest standard concentration value, please dilute the sample appropriately and then re-measure.
- 5) Human anti-mouse and other heterophilic antibodies present in the sample to be tested will interfere with the test results. Please eliminate this factor before testing.
- 6) The test results obtained by other methods are not directly comparable to the test results of this kit.
- 7) Please wear a lab coat and latex gloves for protection during the test. Especially when testing blood or other body fluid samples, please follow the national biological laboratory safety protection regulations.
- 8) Carry out incubation strictly according to the specified time and temperature to ensure accurate results. All reagents must reach room temperature 20-25°C before use. Store reagents refrigerated immediately after use.
- 9) Improper plate washing can lead to inaccurate results. Make sure to absorb as much liquid as possible from the wells before adding substrate. Do not allow the microwells to dry out during incubation.
- 10) Eliminate residual liquid and fingerprints on the bottom of the plate, otherwise it will affect the OD value.
- 11) The substrate chromogenic solution should be colorless or very light in color.
- 12) Avoid cross-contamination of reagents and specimens to avoid erroneous results.
- 13) Avoid direct exposure to strong light during storage and incubation.

- 14) The microplate reader used for detection needs to be equipped with a filter capable of detecting a wavelength of $450\pm 10\text{nm}$, and the optical density range is between 0-3.5. It is recommended to preheat 15 minutes in advance before use.
- 15) The EP tubes and suction tips used in the test are single-use and are strictly prohibited from mixing.

Nanjing BYabsience technology Co.,Ltd

Website: www.byabsience.cn

Official hotline: 025-5229-8998

Supervision phone number:



Sample preparation and storage

The following lists only general guidelines for sample collection and preservation. During the collection and storage of all samples, sodium azide must not be used as a preservative. If the sample is not analyzed immediately, it should be aliquoted and stored frozen, and repeated freezing and thawing should be avoided.

Cell culture supernatant - centrifuge to remove precipitate, analyze immediately or aliquot and store frozen at -20°C.

Serum - Collect blood in a clean test tube, coagulate at room temperature for 30 minutes, centrifuge at 2000×g for 20 minutes, and collect serum. Analyze immediately or aliquot and store frozen at -20°C.

Plasma—anticoagulate with heparin, citrate, or EDTA, and centrifuge at 2000×g for 20 minutes at 2-8°C within 30 minutes of blood draw. To eliminate the effect of platelets, it is recommended to further centrifuge at 10,000 × g for 10 minutes at 2-8°C. Analyze immediately or aliquot and store frozen at -20°C.

Cell lysis buffer - For adherent cells, remove the culture medium and wash with PBS, normal saline or serum-free culture medium. Add an appropriate amount of lysis solution and pipet several times with a gun to fully contact the lysate and cells. Typically after 10 seconds, cells are lysed. For suspended cells, collect the cells by centrifugation and wash them once with PBS, physiological saline or serum-free culture medium. Add an appropriate amount of lysis solution, blow the cells with a gun, and flick them with your fingers to fully lyse the cells. After full lysis, centrifuge at 10000-14000×g for 3-5 minutes and take the supernatant. Analyze immediately or aliquot and store frozen at -20°C.

Tissue homogenate - rinse the tissue with pre-cooled PBS (0.01M, pH=7.4) to remove residual blood (lysed red blood cells in the homogenate will affect the measurement results), weigh and cut the tissue into pieces. Mix the minced tissue with the corresponding volume of PBS (generally

according to the weight to volume ratio of 1:9, for example, 1g of tissue sample corresponds to 9mL of PBS. The specific volume can be adjusted appropriately according to the experimental needs, and recorded. It is recommended to add Protease inhibitor) was added to a glass homogenizer and ground thoroughly on ice. In order to further lyse tissue cells, the homogenate can be sonicated or frozen and thawed repeatedly. Finally, centrifuge the homogenate at 5000 × g for 5 to 10 minutes, and take the supernatant for detection.

Urine - Collect in sterile tubes and centrifuge at 2000×g for 20 minutes. Carefully collect the supernatant. If a precipitate forms, centrifuge again.

Nanjing BYabscience technology Co.,Ltd

Website: www.byabscience.cn

Official hotline: 025-5229-8998

Supervision phone number:



Reagent preparation 1. Before use, all components must be rewarmed for at least 60 minutes to ensure sufficient rewarming to room temperature.

2. Concentrated washing liquid: The concentrated washing liquid taken out from the refrigerator will produce crystals. This is a normal phenomenon. Heating in a water bath will completely dissolve the crystals. Concentrated detergent and distilled water, dilute 1:20, that is, 1 part of concentrated detergent, add 19 parts of distilled water.

Operating procedures: Return all reagents and components to room temperature first. For standards, quality control materials and samples, it is recommended to make duplicate holes.

1. Prepare the working solutions of various components of the kit according to the method described in the previous instructions.
2. Take out the required slats from the aluminum foil bag, seal the remaining slats in a ziplock bag and return it to the refrigerator.
3. Set up standard wells and sample wells, and add 50 μL of standards of different concentrations to each standard well;
4. Add 50 μL of the sample to be tested into the sample well; do not add it to the blank well.
5. Except for the blank well, add 100 μL of horseradish peroxidase (HRP)-labeled detection antibody to each well of the standard well and sample well, seal the reaction well with a sealing film, and keep the temperature in a water bath or constant temperature oven at 37°C. Incubate for 60 minutes.
6. Discard the liquid, pat dry on absorbent paper, fill each well with washing solution (350 μL), let it stand for 1 minute, shake off the washing solution, pat dry on absorbent paper, and repeat washing the plate 5 times (you can also use a plate washer to wash it) plate).

7. Add 50 μ L each of substrates A and B to each well, and incubate at 37°C in the dark for 15 minutes.
8. Add 50 μ L of stop solution to each well, and within 15 minutes, measure the OD value of each well at a wavelength of 450 nm.

[Interpretation of test results]

1. Negative control OD value: less than 0.2.
2. Positive control OD value: greater than 0.8.
3. Positive judgment (Cut-Off value): If the negative control OD value is +0.25, and the sample OD value is greater than the threshold, it is judged as positive, otherwise, it is negative.

Nanjing BYabscience technology Co.,Ltd

Website: www.byabscience.cn

Official hotline: 025-5229-8998

Supervision phone number:



[Problem Analysis] If the experimental results are not good, please take pictures of the color development results in time, save the experimental data, keep the used strips and unused reagents, and then contact our company's technical support to solve the problem for you. At the same time, you can also refer to the following information: [Questions and Answers]

| Problem description | Possible reasons | Corresponding countermeasures Corresponding countermeasures |
|------------------------------------|---|--|
| standard curve gradient difference | Incorrect liquid aspiration or | Check pipettes and tips |
| | Equilibration time is too short | Ensure sufficient balancing time |
| | Incomplete washing | Ensure the washing time and number of washings and the amount of liquid added to each hole |
| Very weak or colorless | Incubation time too short | Ensure adequate incubation time |
| | The experimental temperature is incorrect | Use recommended experimental temperatures |
| | Insufficient reagent volume or missing addition | Check the liquid aspiration and addition process to ensure that all reagents are added in sufficient |
| | Incorrect dilution | |
| | Enzyme label inactivation or substrate failure | Mix enzyme conjugate and substrate and check by rapid color development |
| Reading value is low | Microplate reader settings are incorrect | Check the wavelength and filter |
| | | Turn on the microplate reader and preheat it in advance |
| Large coefficient of variation | Adding fluid incorrectly | Check the filling situation |
| High background value | The working concentration of the | Use the recommended dilution |
| | Incomplete washing of enzyme plate | Ensure that each step of cleaning is complete; if using an automatic plate washer, please check whether all outlets are blocked; |
| | The lotion is contaminated | Prepare fresh lotion |
| Low sensitivity | Improper storage of ELISA kits | Store relevant reagents according to |
| | Not terminated before reading | Stop solution should be added to |

Nanjing BYabscience technology Co.,Ltd

Website: www.byabscience.cn

Official hotline: 025-5229-8998

Supervision phone number:



statement

1. Limited by the existing conditions and scientific and technological level, it is not possible to conduct comprehensive identification and analysis of all raw materials.

This product may have certain quality and technical risks.

2. This kit removes/reduces some endogenous interfering factors in biological samples during the development process. Not all possible influencing factors have been removed.

3. The final experimental results are closely related to factors such as the effectiveness of the reagents, the relevant operations of the experimenter, and the experimental environment at the time. Our company is only responsible for the kit itself and is not responsible for the sample consumption caused by the use of the kit.

Please use Before use, the user should fully consider the possible usage of the sample and reserve sufficient samples.

4. In order to achieve good experimental results, please only use the reagents provided in our company's kits, do not mix products from other manufacturers, and operate in strict accordance with the instructions.

5. Due to incorrect reagent preparation and microplate reader parameter settings during the operation, abnormal results may result. Please read the instructions carefully and adjust the instrument before the experiment.

6. Even if operated by the same personnel, different results may be obtained in two independent experiments. In order to ensure the reproducibility of the results, it is necessary to control every step of the experimental process.

7. The kits will undergo strict quality inspection before shipment. However, due to factors such as transportation conditions, differences in experimental equipment, etc., user test results may be inconsistent with factory data.

8. This kit has not been compared with similar kits from other manufacturers or products using different methods to detect the same target, so inconsistent test results cannot be ruled out.

9. The kit is for research use only. If it is used for clinical diagnosis or any other purpose, our company will not be responsible for any problems arising therefrom, nor will we assume any legal liability.

Nanjing BYabscience technology Co.,Ltd

Website: www.byabscience.cn

Official hotline: 025-5229-8998

Supervision phone number: